

## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-124946

Date Verified: 8/8/2019

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

AECOM Chemist: Jared DeSadier

AECOM ITR: Jeff Aust

Sample Identification #	Date Collected	Date Received	Matrix	Analysis
G0285-19A	6/7/2019	6/8/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
G0111-19A	6/7/2019	6/8/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
G0085-19A	6/7/2019	6/8/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
PZ012-19A	6/7/2019	6/8/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
G0115-19A	6/7/2019	6/8/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
G0095-19A	6/7/2019	6/8/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
G0121-19A	6/7/2019	6/8/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
G0119-19A	6/7/2019	6/8/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
G0101-19A	6/7/2019	6/8/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
G0066R-19A	6/7/2019	6/8/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
G0100-19A	6/7/2019	6/8/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
PZ010-19A	6/7/2019	6/8/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
G0120-19A	6/7/2019	6/8/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
PZ011-19A	6/7/2019	6/8/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)

# CHAAP Data Verification

Laboratory and SDG#: TADenver 280-124946

AECOM Chemist: Jared DeSadier

Date Verified: 8/8/2019

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

Sample Identification #	Date Collected	Date Received	Matrix	Analysis
G0116-19A	6/7/2019	6/8/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
PZ014-19A	6/7/2019	6/8/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
G0117-19A	6/7/2019	6/8/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
PZ013-19A	6/7/2019	6/8/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)

## 1.0 Laboratory Case Narrative \ Cooler Receipt Form

Verification Criteria	Yes	No	N/A
Were any DoD QSM deviations noted in the laboratory case narrative?	X		
Were DoD QSM corrective actions followed if deviations were noted?	X		
Were any issues noted in the cooler receipt form?		X	

The laboratory case narrative indicated that some surrogate, MS/MSD, and LCS/LCSD recoveries were outside evaluation criteria and some samples were analyzed outside of holding time criteria. These issues are discussed further in the ADR report.

RPD between the primary and confirmation column for some explosives samples was above evaluation criteria. This issue is discussed further in Section 10.0. Some explosives CCV %Ds were outside of evaluation criteria. This issue is discussed further in Section 5.0. A discrepancy was noted between the COC and sample label for the collection time of sample. This issue is discussed further in Section 2.0.

No other issues were noted in the case narrative or cooler receipt form.

## 2.0 Sample Documentation

Verification Criteria	Yes	No
Were all samples documented correctly on the chain-of-custody (COC) and samples labels?		X
Were all sample identifications (IDs) documented correctly on sample labels?		X
Did samples listed on COCs match the sample labels?		X
Were samples relinquished properly on the COC?	X	

## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-124946

AECOM Chemist: Jared DeSadier

Date Verified: 8/8/2019

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

A discrepancy was noted between the COC and sample label for the collection time of sample. Per the AECOM chemist, the sample was logged via the COC and no qualification of data was required.

### 3.0 Initial Calibration

Method 8330A Initial Calibration Criteria			
<b>Instrument:</b>	CHHPL G2 LUNA		
<b>Date of Calibration:</b>	5/7/2019		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was at least a five point calibration completed for all analytes prior to sample analysis and one option below?	X		
Option 1: RSD for each analyte $\leq 20\%$ ?	X		
Option 2: If linear least squares regression was used was the $r^2 \geq 0.99$ ?			X
Option 3: If non-linear regression was used was the coefficient of determination $r^2 \geq 0.99$ ?			X
If non-linear regression was used were 6 points used for second order and 7 points for third order?			X

Method 8330A Initial Calibration Criteria			
<b>Instrument:</b>	CHHPLC X3		
<b>Date of Calibration:</b>	5/14/2019		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was at least a five point calibration completed for all analytes prior to sample analysis and one option below?	X		
Option 1: RSD for each analyte $\leq 20\%$ ?	X		
Option 2: If linear least squares regression was used was the $r^2 \geq 0.99$ ?			X
Option 3: If non-linear regression was used was the coefficient of determination $r^2 \geq 0.99$ ?			X
If non-linear regression was used were 6 points used for second order and 7 points for third order?			X

Method 8330A Initial Calibration Criteria			
<b>Instrument:</b>	CHHPLC X3		
<b>Date of Calibration:</b>	6/20/2019		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was at least a five point calibration completed for all analytes prior to sample analysis and one option below?	X		
Option 1: RSD for each analyte $\leq 20\%$ ?	X		
Option 2: If linear least squares regression was used was the $r^2 \geq 0.99$ ?			X
Option 3: If non-linear regression was used was the coefficient of determination $r^2 \geq 0.99$ ?			X
If non-linear regression was used were 6 points used for second order and 7 points for third order?			X

## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-124946

AECOM Chemist: Jared DeSadier

Date Verified: 8/8/2019

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

Method RSK-175 Initial Calibration Criteria			
<b>Instrument:</b>	VGC J		
<b>Date of Calibration:</b>	04/15/2019		
	Yes	No	N/A
Was at least a five point calibration completed for all analytes prior to sample analysis and one option below?	X		
Option 1: RSD for each analyte $\leq 25\%$ ?	X		
Option 2: If linear least squares regression was used was the $r^2 \geq 0.99$ ?	X		
Option 3: If non-linear regression was used was the coefficient of determination $r^2 \geq 0.99$ ?			X
If non-linear regression was used were 6 points used for second order and 7 points for third order?			X

Method 9056A Initial Calibration Criteria			
<b>Instrument:</b>	WC IonChrom10		
<b>Date of Calibration:</b>	7/5/2019		
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \geq 0.99$ ?	X		

Method 9056A Initial Calibration Criteria			
<b>Instrument:</b>	WC IonChrom8		
<b>Date of Calibration:</b>	6/3/2019		
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \geq 0.99$ ?	X		

Method 350.1 Initial Calibration Criteria			
<b>Instrument:</b>	WC Alp 3		
<b>Date of Calibration:</b>	6/21/2019		
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \geq 0.99$ ?	X		

Method 353.2 Initial Calibration Criteria			
<b>Instrument:</b>	WC Alp 2		
<b>Date of Calibration:</b>	6/27/2019		
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \geq 0.99$ ?	X		

## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-124946

AECOM Chemist: Jared DeSadier

Date Verified: 8/8/2019

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

Method 353.2 Initial Calibration Criteria			
Instrument:	WC Alp 2		
Date of Calibration:	7/1/2019		
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \geq 0.99$ ?	X		

Method 351.2 Initial Calibration Criteria			
Instrument:	WC Astoria		
Date of Calibration:	6/26/2019		
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \geq 0.99$ ?	X		

Method 9060A Initial Calibration Criteria			
Instrument:	WC SHI3		
Date of Calibration:	7/3/2019		
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \geq 0.99$ ?	X		

Method 9060A Initial Calibration Criteria			
Instrument:	WC SHI3		
Date of Calibration:	7/5/2019		
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \geq 0.99$ ?	X		

### 4.0 Initial Calibration Verification [(ICV) Second Source]

Method 8330A ICV Criteria (Filename)			
	05070015.D		
Instrument:	CHHLPC G2 LUNA		
Date of Initial Calibration Verification:	5/7/2019		
	Yes	No	N/A
Was the ICV analyzed after each calibration?	X		
Was the ICV for all analytes within $\pm 15\%$ of the true value?	X		

Method 8330A ICV Criteria (Filename)			
	05080007.D		
Instrument:	CHHLPC G2 LUNA		
Date of Initial Calibration Verification:	5/8/2019		
	Yes	No	N/A
Was the ICV analyzed after each calibration?	X		
Was the ICV for all analytes within $\pm 15\%$ of the true value?	X		

## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-124946

AECOM Chemist: Jared DeSadier

Date Verified: 8/8/2019

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

<b>Method 8330A ICV Criteria (Filename)</b>	<b>0514B015.D</b>		
<b>Instrument:</b>	<b>CHHPLC X3</b>		
<b>Date of Initial Calibration Verification:</b>	<b>5/14/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the ICV analyzed after each calibration?	X		
Was the ICV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A ICV Criteria (Filename)</b>	<b>0514B033.D</b>		
<b>Instrument:</b>	<b>CHHPLC X3</b>		
<b>Date of Initial Calibration Verification:</b>	<b>5/15/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the ICV analyzed after each calibration?	X		
Was the ICV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A ICV Criteria (Filename)</b>	<b>06200015.D</b>		
<b>Instrument:</b>	<b>CHHPLC X3</b>		
<b>Date of Initial Calibration Verification:</b>	<b>6/20/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the ICV analyzed after each calibration?	X		
Was the ICV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A ICV Criteria (Filename)</b>	<b>06200033.D</b>		
<b>Instrument:</b>	<b>CHHPLC X3</b>		
<b>Date of Initial Calibration Verification:</b>	<b>6/21/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the ICV analyzed after each calibration?	X		
Was the ICV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method RSK-175 ICV Criteria (Filename)</b>	<b>04151911.D</b>		
<b>Instrument:</b>	<b>VGC J</b>		
<b>Date of Initial Calibration Verification:</b>	<b>7/2/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the ICV analyzed after each calibration?	X		
Was the ICV for all analytes within $\pm 25\%$ of the true value?	X		

<b>Method 9056A ICV</b>	<b>WC IonChrom10</b>		
<b>Date of Initial Calibration Verification:</b>	<b>7/5/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm 10\%$ of the true value?	X		

## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-124946

AECOM Chemist: Jared DeSadier

Date Verified: 8/8/2019

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

Method 9056A ICV	WC IonChrom8		
Date of Initial Calibration Verification:	6/3/2019		
	Yes	No	N/A
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm 10\%$ of the true value?	X		

Method 350.1 ICV Criteria	10:12		
Instrument:	WC Alp 3		
Date of Initial Calibration Verification:	6/21/2019		
	Yes	No	N/A
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm 10\%$ of the true value?	X		

Method 353.2 ICV Criteria (Filename)	WC Alp 2		
Date of Initial Calibration Verification:	7/1/2019		
	Yes	No	N/A
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm 10\%$ of the true value?	X		

Method 351.2 ICV Criteria (Filename)	18:05		
Instrument:	WC Astoria		
Date of Initial Calibration Verification:	6/26/2019		
	Yes	No	N/A
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm 10\%$ of the true value?	X		

Method 9060A ICV Criteria (Filename)	15:51		
Instrument:	WC SHI3		
Date of Initial Calibration Verification:	7/3/2019		
	Yes	No	N/A
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm 10\%$ of the true value?	X		

## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-124946

AECOM Chemist: Jared DeSadier

Date Verified: 8/8/2019

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

<b>Method 9060A ICV Criteria (Filename)</b>	<b>11:27</b>		
<b>Instrument:</b>	<b>WC_SHI3</b>		
<b>Date of Initial Calibration Verification:</b>	<b>7/5/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm 10\%$ of the true value?	X		

### 5.0 Continuing Calibration Verification (CCV)

<b>Method 8330A CCV Criteria (Filename)</b>	<b>06190007_8.D</b>		
<b>Instrument:</b>	<b>CHHPLC_G2_LUNA</b>		
<b>Date of Calibration Verification:</b>	<b>6/19/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A CCV Criteria (Filename)</b>	<b>06190019_20.D</b>		
<b>Instrument:</b>	<b>CHHPLC_G2_LUNA</b>		
<b>Date of Calibration Verification:</b>	<b>6/20/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A CCV Criteria (Filename)</b>	<b>06190031_2.D</b>		
<b>Instrument:</b>	<b>CHHPLC_G2_LUNA</b>		
<b>Date of Calibration Verification:</b>	<b>6/20/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		



## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-124946

AECOM Chemist: Jared DeSadier

Date Verified: 8/8/2019

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

Method 8330A CCV Criteria (Filename)	06190043_4.D		
Instrument:	CHHPLC_G2_LUNA		
Date of Calibration Verification:	6/20/2019		
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

The %D for tetryl was outside evaluation criteria on the confirmation column. The sample was not being confirmed for tetryl and no qualification of data was required.

Method 8330A CCV Criteria (Filename)	06190051_2.D		
Instrument:	CHHPLC_G2_LUNA		
Date of Calibration Verification:	6/20/2019		
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

Method 8330A CCV Criteria (Filename)	0621007_8.D		
Instrument:	CHHPLC_G2_LUNA		
Date of Calibration Verification:	6/21/2019		
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

Method 8330A CCV Criteria (Filename)	06210011_2.D		
Instrument:	CHHPLC_G2_LUNA		
Date of Calibration Verification:	6/21/2019		
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-124946

AECOM Chemist: Jared DeSadier

Date Verified: 8/8/2019

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

<b>Method 8330A CCV Criteria (Filename)</b>	<b>06170041_3.D</b>		
<b>Instrument:</b>	<b>CHHPLC_X3</b>		
<b>Date of Calibration Verification:</b>	<b>6/18/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A CCV Criteria (Filename)</b>	<b>06170054_6.D</b>		
<b>Instrument:</b>	<b>CHHPLC_X3</b>		
<b>Date of Calibration Verification:</b>	<b>6/18/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

The CCV %Ds for nitrobenzene, 2-nitrotoluene, 3-nitrotoluene, and 4-nitrotoluene were outside evaluation criteria on the confirmation column. The samples were not being confirmed for these analytes and no qualification of data was required.

<b>Method 8330A CCV Criteria (Filename)</b>	<b>06190069_71.D</b>		
<b>Instrument:</b>	<b>CHHPLC_X3</b>		
<b>Date of Calibration Verification:</b>	<b>6/20/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A CCV Criteria (Filename)</b>	<b>007-7701_045-7901.D</b>		
<b>Instrument:</b>	<b>CHHPLC_X3</b>		
<b>Date of Calibration Verification:</b>	<b>6/20/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-124946

AECOM Chemist: Jared DeSadier

Date Verified: 8/8/2019

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

<b>Method 8330A CCV Criteria (Filename)</b>	<b>06250018_20.D</b>		
<b>Instrument:</b>	<b>CHHPLC X3</b>		
<b>Date of Calibration Verification:</b>	<b>6/25/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A CCV Criteria (Filename)</b>	<b>06250026_8.D</b>		
<b>Instrument:</b>	<b>CHHPLC X3</b>		
<b>Date of Calibration Verification:</b>	<b>6/25/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A CCV Criteria (Filename)</b>	<b>06250026_8.D</b>		
<b>Instrument:</b>	<b>CHHPLC X3</b>		
<b>Date of Calibration Verification:</b>	<b>6/25/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A CCV Criteria (Filename)</b>	<b>06250038_40.D</b>		
<b>Instrument:</b>	<b>CHHPLC X3</b>		
<b>Date of Calibration Verification:</b>	<b>6/26/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-124946

AECOM Chemist: Jared DeSadier

Date Verified: 8/8/2019

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

<b>Method RSK-175 CCV Criteria (Filename)</b>	<b>06131928.D</b>		
<b>Instrument:</b>	<b>VGC J</b>		
<b>Date of Calibration Verification:</b>	<b>6/13/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV %D or %drift for all target compounds $\leq$ 25%?	X		

<b>Method RSK-175 CCV Criteria (Filename)</b>	<b>06131942.D</b>		
<b>Instrument:</b>	<b>VGC J</b>		
<b>Date of Calibration Verification:</b>	<b>6/13/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm$ 25% of the true value?	X		

<b>Method RSK-175 CCV Criteria (Filename)</b>	<b>06131954.D</b>		
<b>Instrument:</b>	<b>VGC J</b>		
<b>Date of Calibration Verification:</b>	<b>6/13/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm$ 25% of the true value?	X		

<b>Method RSK-175 CCV Criteria (Filename)</b>	<b>06141932.D</b>		
<b>Instrument:</b>	<b>VGC J</b>		
<b>Date of Calibration Verification:</b>	<b>6/14/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm$ 25% of the true value?	X		

## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-124946

AECOM Chemist: Jared DeSadier

Date Verified: 8/8/2019

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

<b>Method RSK-175 CCV Criteria (Filename)</b>	<b>06141947.D</b>		
<b>Instrument:</b>	<b>VGC J</b>		
<b>Date of Calibration Verification:</b>	<b>6/14/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 25\%$ of the true value?	X		

<b>Method RSK-175 CCV Criteria (Filename)</b>	<b>06171925.D</b>		
<b>Instrument:</b>	<b>VGC J</b>		
<b>Date of Calibration Verification:</b>	<b>6/17/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 25\%$ of the true value?	X		

<b>Method RSK-175 CCV Criteria (Filename)</b>	<b>044F4501.D</b>		
<b>Instrument:</b>	<b>VGC J</b>		
<b>Date of Calibration Verification:</b>	<b>6/17/2019</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 25\%$ of the true value?	X		

<b>Method 9056A, Instrument: WC IonChrom10, All CCVs on 7/5/2019</b>	<b>Yes</b>	<b>No</b>
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X	
Were the CCVs for all analytes within $\pm 10\%$ of the true value?	X	

<b>Method 9056A, Instrument: WC IonChrom8, All CCVs on 6/3/2019</b>	<b>Yes</b>	<b>No</b>
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X	
Were the CCVs for all analytes within $\pm 10\%$ of the true value?	X	

<b>Method 9056A, Instrument: WC IonChrom8, All CCVs on 7/16/2019</b>	<b>Yes</b>	<b>No</b>
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X	
Were the CCVs for all analytes within $\pm 10\%$ of the true value?	X	

<b>Method 350.1, Instrument: WC Alp 3, All CCVs on 6/21/2019</b>	<b>Yes</b>	<b>No</b>
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X	
Were the CCVs for all analytes within $\pm 10\%$ of the true value?	X	

# CHAAP Data Verification

Laboratory and SDG#: TADenver 280-124946

AECOM Chemist: Jared DeSadier

Date Verified: 8/8/2019

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

Method 353.2, Instrument: WC Alp 2, All CCVs on 6/27/2019	Yes	No
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X	
Were the CCVs for all analytes within $\pm 10\%$ of the true value?	X	

Method 353.2, Instrument: WC Alp 2, All CCVs on 7/1/2019	Yes	No
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X	
Were the CCVs for all analytes within $\pm 10\%$ of the true value?	X	

Method 351.2, Instrument: WC Astoria, All CCVs on 6/26/2019	Yes	No
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X	
Were the CCVs for all analytes within $\pm 10\%$ of the true value?	X	

Method 9060A, Instrument: WC SHI3, All CCVs on 7/3/2019	Yes	No
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X	
Were the CCVs for all analytes within $\pm 10\%$ of the true value?	X	

Method 9060A, Instrument: WC SHI3, All CCVs on 7/5/2019	Yes	No
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X	
Were the CCVs for all analytes within $\pm 10\%$ of the true value?	X	

## 6.0 Blank Samples

Blank Criteria	Yes	No	N/A
Were method blanks analyzed with every preparatory batch?	X		
Were target analytes detected $> \frac{1}{2}$ the LOQ and $> 1/10$ the amount measured in any sample or $1/10$ the regulatory limit (whichever is greater)?		X	
Were target analytes detected in method, trip or calibration blanks?		X	

## 7.0 Field Duplicate Samples

Field Duplicate Criteria	Yes	No	N/A
Were field duplicate samples collected for this SDG? (if yes, list below)	X		
Were parent sample / field duplicate RPDs $\leq 30\%$ for water samples and $\leq 50\%$ for soils for analytes that had concentrations $> 5x$ the LOQ?	X		
Were the differences between the parent sample / field duplicate $< 2x$ the LOQ for analytes that had concentrations $< 5x$ the LOQ?	X		

Parent ID	Duplicate ID
G0085-19A	G0285-19A

## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-124946

AECOM Chemist: Jared DeSadier

Date Verified: 8/8/2019

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

### 8.0 Sensitivity

Sensitivity Criteria	Yes	No	N/A
Was the laboratory sensitivity consistent with project (QAPP) requirements?	X		
Did all analytes meet sensitivity requirements?	X		

### 9.0 Additional Qualifications

Additional Qualification Criteria	Yes	No	N/A
Were common laboratory contaminants detected?		X	
Was professional judgment used to qualify data (if yes, list below)?	X		

The RPD between the primary and confirmation column for some explosives samples was above evaluation criteria. Qualification of data is shown in the table below.

Sample ID	Analysis	Analyte	RPD	Qual
G0095-19A	Explosives	RDX	108.9	J
G0121-19A	Explosives	HMX	44.8	J
G0121-19A	Explosives	MNX	187.2	J
G0121-19A	Explosives	4-Amino-2,6-dinitrotoluene	147.4	J
G0111-19A	Explosives	HMX	198.4	J
G0085-19A	Explosives	HMX	51.5	J
G0085-19A	Explosives	Tetryl	138.1	J
G0066R-19A	Explosives	2-amino-4,6-dinitrotoluene	42.5	J
G0117-19A	Explosives	MNX	188.1	J
PZ013-19A	Explosives	Tetryl	190.9	J
PZ010-19A	Explosives	Tetryl	142.5	J
PZ011-19A	Explosives	Tetryl	197.3	J

### 10.0 Completeness

Completeness Criteria	Yes	No	N/A
Were any data rejected during the verification process?		X	
Were any samples lost, broken, or in any other manner in not verified?		X	
Were requested sample analyses performed, the correct analyte lists used, and correct sample preparation and analyses methods and units utilized?	X		