Laboratory and SDG#: TADenver 280-124912 AECOM Chemist: Jared DeSadier

Date Verified: 8/1/2019 AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018) Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

Sample Identification #	Date Collected	Date Received	Matrix	Analysis
PZ005-19A	6/5/2019	6/6/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
G0103-19A	6/5/2019	6/6/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
G0104-19A	6/5/2019	6/6/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
G0102-19A	6/5/2019	6/6/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
PZ001-19A	6/5/2019	6/6/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
PZ004-19A	6/5/2019	6/6/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
G0044-19A	6/5/2019	6/6/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
PZ015-19A	6/5/2019	6/6/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
G0049-19A	6/5/2019	6/6/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
G0048-19A	6/5/2019	6/6/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
G0023-19A	6/5/2019	6/6/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)
PZ007-19A	6/5/2019	6/6/2019	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), MEE (RSK-175), DOC (9060A), Sulfate (9056A) Sulfide (9034), Alkalinity (2320B)

#### 1.0 Laboratory Case Narrative \ Cooler Receipt Form

Verification Criteria	Yes	No	N/A
Were any DoD QSM deviations noted in the laboratory case narrative?	X		
Were DoD QSM corrective actions followed if deviations were noted?	X		
Were any issues noted in the cooler receipt form?		X	

Laboratory and SDG#: TADenver 280-124912 AECOM Chemist: Jared DeSadier

Date Verified: 8/1/2019 AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018) Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

The laboratory case narrative indicated that some surrogate, MS/MSD, and LCS/LCSD recoveries were outside evaluation criteria These issues are discussed further in the ADR report.

DOC and sulfate were detected in some blank samples. This issue is discussed further in Section 6.0. RPD between the primary and confirmation column for some explosives samples was above evaluation criteria. This issue is discussed further in Section 10.0. An explosives CCV %D was outside of evaluation criteria. This issue is discussed further in Section 5.0.

No other issues were noted in the case narrative or cooler receipt form.

#### 2.0 Sample Documentation

Verification Criteria	Yes	No
Were all samples documented correctly on the chain-of-custody (COC) and samples labels?	X	
Were all sample identifications (IDs) documented correctly on sample labels?		
Did samples listed on COCs match the sample labels?		
Were samples relinquished properly on the COC?	X	

#### 3.0 Initial Calibration

Method 8330A Initial Calibration Criteria				
Instrument:		CHHPL_G2_LUNA		
Date of Calibration:	5/7/2019			
	Yes	No	N/A	
Was at least a five point calibration completed for all analytes prior to sample analysis and one option below?	X			
Option 1: RSD for each analyte ≤ 20%?	X			
Option 2: If linear least squares regression was used was the $r^2 \ge 0.99$ ?			X	
Option 3: If non-linear regression was used was the coefficient of determination $r^2 \ge 0.99$ ?			X	
If non-linear regression was used were 6 points used for second order and 7 points for third order?			X	

Laboratory and SDG#: TADenver 280-124912 AECOM Chemist: Jared DeSadier

Date Verified: 8/1/2019 AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Method 8330A Initial Calibration Criteria				
Instrument:		CHHPLC_X3		
Date of Calibration:	5/14/2019			
	Yes	No	N/A	
Was at least a five point calibration completed for all analytes prior to sample analysis and one option below?	X			
Option 1: RSD for each analyte $\leq 20\%$ ?	X			
Option 2: If linear least squares regression was used was the $r^2 \ge 0.99$ ?			X	
Option 3: If non-linear regression was used was the coefficient of determination $r^2 \ge 0.99$ ?			X	
If non-linear regression was used were 6 points used for second order and 7 points for third order?			X	

Method 8330A Initial Calibration Criteria					
Instrument:		CHHPLC_X3			
Date of Calibration:		7/1/2019			
	Yes	No	N/A		
Was at least a five point calibration completed for all analytes prior to sample analysis and one option below?	X				
Option 1: RSD for each analyte ≤ 20%?	X				
Option 2: If linear least squares regression was used was the $r^2 \ge 0.99$ ?			X		
Option 3: If non-linear regression was used was the coefficient of determination $r^2 \ge 0.99$ ?			X		
If non-linear regression was used were 6 points used for second order and 7 points for third order?			X		

Method RSK-175 Initial Calibration Criteria					
Instrument:		VGC_J			
Date of Calibration:		04/15/2019			
	Yes	No	N/A		
Was at least a five point calibration completed for all analytes prior to sample analysis and one option below?	X				
Option 1: RSD for each analyte $\leq 25\%$ ?	X				
Option 2: If linear least squares regression was used was the $r^2 \ge 0.99$ ?	X				
Option 3: If non-linear regression was used was the coefficient of determination $r^2 \ge 0.99$ ?			X		
If non-linear regression was used were 6 points used for second order and 7 points for third order?			X		

Method 9056A Initial Calibration Criteria			
Instrument:			om11
Date of Calibration:		6/26/2019	
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \ge 0.99$ ?	X		

Laboratory and SDG#: TADenver 280-124912 AECOM Chemist: Jared DeSadier

Date Verified: 8/1/2019 AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Method 350.1 Initial Calibration Criteria			
Instrument:			3
Date of Calibration:		6/17/2019	
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \ge 0.99$ ?	X		

Method 350.1 Initial Calibration Criteria					
Instrument:		WC_Alp 3			
Date of Calibration:		6/20/2019			
	Yes	No	N/A		
Was a minimum of three standards and a calibration blank used for ICAL?	X				
Was $r^2 \ge 0.99$ ?	X				

Method 353.2 Initial Calibration Criteria				
Instrument:		WC_Alp 2		
Date of Calibration:		6/26/2019		
	Yes	No	N/A	
Was a minimum of three standards and a calibration blank used for ICAL?	X			
Was $r^2 \ge 0.99$ ?	X			

Method 351.2 Initial Calibration Criteria			
Instrument:		WC_Astoria	
Date of Calibration:		6/25/2019	
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \ge 0.99$ ?	X		

Method 9060A Initial Calibration Criteria			
Instrument:	W	C_SH	<b>I3</b>
Date of Calibration:		7/1/2019	
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \ge 0.99$ ?	X		

Laboratory and SDG#: TADenver 280-124912 AECOM Chemist: Jared DeSadier

Date Verified: 8/1/2019 AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)
Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

#### 4.0 Initial Calibration Verification [(ICV) Second Source]

Method 8330A ICV Criteria (Filename)	05070015.D		
Instrument:	CHHLI	PC_G2_	LUNA
Date of Initial Calibration Verification:	5/7/2019		
	Yes	No	N/A
Was the ICV analyzed after each calibration?	X		
Was the ICV for all analytes within ± 15% of the true value?	X		

Method 8330A ICV Criteria (Filename)	05080007.D			
Instrument:	CHHLI	PC_G2_	LUNA	
Date of Initial Calibration Verification:	5	5/8/2019		
	Yes	No	N/A	
Was the ICV analyzed after each calibration?	X			
Was the ICV for all analytes within $\pm$ 15% of the true value?	X			

Method 8330A ICV Criteria (Filename)	05	0514B015.D		
Instrument:	CE	HPLC	X3	
Date of Initial Calibration Verification:	45	5/14/2019		
	Yes	No	N/A	
Was the ICV analyzed after each calibration?	X			
Was the ICV for all analytes within $\pm$ 15% of the true value?	X			

Method 8330A ICV Criteria (Filename)	0	0514B033.D		
Instrument:	C	CHHPLC_X3		
Date of Initial Calibration Verification:		5/15/2019		
	Yes	No	N/A	
Was the ICV analyzed after each calibration?	X			
Was the ICV for all analytes within $\pm$ 15% of the true value?	3.7			

Method 8330A ICV Criteria (Filename)	0′	5.D		
Instrument:	CHHPLC_X3		_X3	
Date of Initial Calibration Verification:		7/1/2019		
	Yes	No	N/A	
Was the ICV analyzed after each calibration?	X			
Was the ICV for all analytes within $\pm$ 15% of the true value?	X			

Laboratory and SDG#: TADenver 280-124912 AECOM Chemist: Jared DeSadier

Date Verified: 8/1/2019 AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Method 8330A ICV Criteria (Filename)	0701033.D		.D
Instrument:	CH	CHHPLC_X3	
Date of Initial Calibration Verification:		6/21/2019	
	Yes No N/A		N/A
Was the ICV analyzed after each calibration?	X		
Was the ICV for all analytes within $\pm$ 15% of the true value?	X		

Method RSK-175 ICV Criteria (Filename)	04	.D	
Instrument:	VGC_J		Г
Date of Initial Calibration Verification:	4/15/2019		
	Yes No N/A		
Was the ICV analyzed after each calibration?	X		
Was the ICV for all analytes within $\pm$ 25% of the true value?	X		

Method 9056A ICV	WC_IonChrom11		
Date of Initial Calibration Verification:	6/	6/26/2019	
	Yes	No	N/A
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm$ 10% of the true value?	X		

Method 350.1 ICV Criteria		11:32	
Instrument:	W	WC_Alp 3	
Date of Initial Calibration Verification:	6.	6/17/2019	
	Yes	No	N/A
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm$ 10% of the true value?	X		

Method 350.1 ICV Criteria		11:59	
Instrument:	W	WC_Alp 3	
Date of Initial Calibration Verification:	6	6/20/2019	
	Yes	No	N/A
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm$ 10% of the true value?	X		

Method 353.2 ICV Criteria (Filename)	WC_Alp 2		
Date of Initial Calibration Verification:	6/	6/26/2019	
	Yes	No	N/A
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm$ 10% of the true value?	X		·

Laboratory and SDG#: TADenver 280-124912 AECOM Chemist: Jared DeSadier

Date Verified: 8/1/2019 AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018) Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

Method 351.2 ICV Criteria (Filename)		19:01	
Instrument:	W	WC_Astoria	
Date of Initial Calibration Verification:	6.	6/25/2019	
	Yes	No	N/A
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm$ 10% of the true value?	X		

Method 9060A ICV Criteria (Filename)	14:01		
Instrument:	WC_SHI3		I3
Date of Initial Calibration Verification:	7/1/2019		)
	Yes	No	N/A
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm$ 10% of the true value?	X		

#### 5.0 Continuing Calibration Verification (CCV)

Method 8330A CCV Criteria (Filename)	06140040_1.D		1.D
Instrument:	CHHPI	CHHPLC_G2_LUNA	
Date of Calibration Verification:	6	6/14/2019	
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm$ 15% of the true value?		X	

The %D for RDX (60.5) was outside evaluation criteria. The RFs indicated a high bias. Qualification of data is shown in the table below.

Sample ID	Parameter	Analyte	Qual
G0044-19A	Explosives	RDX	J
PZ015-19A	Explosives	RDX	J

Laboratory and SDG#: TADenver 280-124912 AECOM Chemist: Jared DeSadier

Date Verified: 8/1/2019 AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Method 8330A CCV Criteria (Filename)	06140049_50.D		50.D
Instrument:	CHHP	CHHPLC_G2_LUNA	
Date of Calibration Verification:	6	6/15/2019	
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm$ 15% of the true value?	X		

Method 8330A CCV Criteria (Filename)	061	06170025_6.D	
Instrument:	CHHP	CHHPLC_G2_LUNA	
Date of Calibration Verification:	6	6/18/2019	
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm$ 15% of the true value?	X		

Method 8330A CCV Criteria (Filename)	0617003	06170038_007-3801.D	
Instrument:	CHHPI	CHHPLC_G2_LUNA	
Date of Calibration Verification:	6/18/2019		9
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm$ 15% of the true value?	X		

Method 8330A CCV Criteria (Filename)	06130036_8.D		8.D
Instrument:	СН	CHHPLC X3	
Date of Calibration Verification:	6	6/13/2019	
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm$ 15% of the true value?	X		

Laboratory and SDG#: TADenver 280-124912 AECOM Chemist: Jared DeSadier

Date Verified: 8/1/2019 AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Method 8330A CCV Criteria (Filename)	0613	06130049_51.D	
Instrument:	СН	CHHPLC_X3	
Date of Calibration Verification:	6	6/14/2019	
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm$ 15% of the true value?	X		

Method 8330A CCV Criteria (Filename)	06130062_4.D		4.D
Instrument:	СН	CHHPLC_X3	
Date of Calibration Verification:	6	6/14/2019	
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm$ 15% of the true value?	X		

Method 8330A CCV Criteria (Filename)	007-71	007-7101_038-7301.	
Instrument:	СН	CHHPLC X3	
Date of Calibration Verification:	6	6/14/2019	
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm$ 15% of the true value?	X		

Method 8330A CCV Criteria (Filename)	06140066_8.D		8.D
Instrument:	СН	CHHPLC_X3	
Date of Calibration Verification:	6	6/15/2019	
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm$ 15% of the true value?	X		

Laboratory and SDG#: TADenver 280-124912 AECOM Chemist: Jared DeSadier

Date Verified: 8/1/2019 AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Method 8330A CCV Criteria (Filename)	061	06140079_81.D	
Instrument:	СН	CHHPLC_X3	
Date of Calibration Verification:	6	6/15/2019	
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm$ 15% of the true value?	X		

Method 8330A CCV Criteria (Filename)	071	07110042_4.D	
Instrument:	СН	CHHPLC X3	
Date of Calibration Verification:	7	7/11/2019	
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm$ 15% of the true value?	X		

Method 8330A CCV Criteria (Filename)	071	07110055_7.D		
Instrument:	CH	CHHPLC_X3		
Date of Calibration Verification:	7	7/12/2019		
	Yes	No	N/A	
Was the CCV analyzed daily before sample analysis?	X			
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X			
Was the CCV for all analytes within $\pm$ 15% of the true value?	X			

Method 8330A CCV Criteria (Filename)	071	07110068_70.D	
Instrument:	СН	CHHPLC X3	
Date of Calibration Verification:	7	7/12/2019	
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm$ 15% of the true value?	X		

Laboratory and SDG#: TADenver 280-124912 AECOM Chemist: Jared DeSadier

Date Verified: 8/1/2019 AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Method 8330A CCV Criteria (Filename)	071	07150007_9.D	
Instrument:	СН	CHHPLC X3	
Date of Calibration Verification:	7	7/15/2019	
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm$ 15% of the true value?	X		

Method 8330A CCV Criteria (Filename)	071	07150020_22.D	
Instrument:	CH	CHHPLC X3	
Date of Calibration Verification:	7	7/15/2019	
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm$ 15% of the true value?	X		

Method RSK-175 CCV Criteria (Filename)	0	06111928.D	
Instrument:		VGC J	
Date of Calibration Verification:		6/11/2019	
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV %D or %drift for all target compounds ≤ 20%?	X		

Method RSK-175 CCV Criteria (Filename)	0	06111947.D	
Instrument:		VGC J	
Date of Calibration Verification:		6/11/2019	
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm$ 25% of the true value?	X		

Laboratory and SDG#: TADenver 280-124912 AECOM Chemist: Jared DeSadier

Date Verified: 8/1/2019 AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Method RSK-175 CCV Criteria (Filename)	0	06111961.D	
Instrument:		VGC J	
Date of Calibration Verification:		6/11/2019	
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm$ 25% of the true value?	X		

Method RSK-175 CCVRT Criteria (Filename)	0	06121901.D	
Instrument:		VGC_J	
Date of Calibration Verification:		6/12/2019	
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm$ 25% of the true value?	X		

Method RSK-175 CCV Criteria (Filename)	0	06121916.D	
Instrument:		VGC J	
Date of Calibration Verification:		6/12/2019	
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm$ 25% of the true value?	X		

Method 9056A, Instrument: WC_IonChrom11, All CCVs on 6/26/2019	Yes	No
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X	
Were the CCVs for all analytes within $\pm$ 10% of the true value?	X	

Method 9056A, Instrument: WC_IonChrom11, All CCVs on 7/1/2019			
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X		
Were the CCVs for all analytes within $\pm$ 10% of the true value?	X		

Method 350.1, Instrument: WC_Alp 3, All CCVs on 6/17/2019		No
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X	
Were the CCVs for all analytes within $\pm$ 10% of the true value?	X	

Method 350.1, Instrument: WC_Alp 3, All CCVs on 6/20/2019	Yes	No
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X	
Were the CCVs for all analytes within $\pm$ 10% of the true value?	X	

Laboratory and SDG#: TADenver 280-124912 AECOM Chemist: Jared DeSadier

Date Verified: 8/1/2019 AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018) Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

Method 353.2, Instrument: WC_Alp 2, All CCVs on 6/26/2019		
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X	
Were the CCVs for all analytes within $\pm$ 10% of the true value?	X	

Method 351.2, Instrument: WC_Astoria, All CCVs on 6/25/2019	Yes	No
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X	
Were the CCVs for all analytes within $\pm$ 10% of the true value?	X	

Method 9060A, Instrument: WC_SHI3, All CCVs on 7/1/2019			
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X		
Were the CCVs for all analytes within $\pm$ 10% of the true value?	X		

Method 9060A, Instrument: WC_SHI3, All CCVs on 7/2/2019		
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X	
Were the CCVs for all analytes within $\pm$ 10% of the true value?	X	

#### 6.0 Blank Samples

Blank Criteria	Yes	No	N/A
Were method blanks analyzed with every preparatory batch?	X		
Were target analytes detected $> \frac{1}{2}$ the LOQ and $> 1/10$ the amount measured in any sample or $1/10$ the regulatory limit (whichever is greater)?		X	
Were target analytes detected in method, trip or calibration blanks?	X		

Blank ID	Parameter	Analyte	Concentration	LOQ	Units
MB 280-463267/2-A	DOC	DOC	0.177	1.0	ug/L
MB 280-463396/2-A	DOC	DOC	0.183	1.0	ug/L

All analytical data were reported nondetect or at concentrations greater than five times (5X) the associated blank concentration and did not require qualification.

#### 7.0 Field Duplicate Samples

Field Duplicate Criteria	Yes	No	N/A
Were field duplicate samples collected for this SDG? (if yes, list below)		X	
Were parent sample / field duplicate RPDs $\leq$ 30% for water samples and $\leq$ 50% for soils for analytes that had concentrations $>$ 5x the LOQ?			X
Were the differences between the parent sample / field duplicate < 2x the LOQ for analytes that had concentrations < 5x the LOQ?			X

Laboratory and SDG#: TADenver 280-124912 AECOM Chemist: Jared DeSadier

Date Verified: 8/1/2019 AECOM ITR: Jeff Aust

**Guidance: DoD QSM Version 5.1 (January 2017)** 

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018) Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

#### 8.0 Sensitivity

Sensitivity Criteria  Was the leherstowy consistivity consistent with project (OADD) requirements?		No	N/A
Was the laboratory sensitivity consistent with project (QAPP) requirements?			
Did all analytes meet sensitivity requirements?	X		

#### 9.0 Additional Qualifications

Additional Qualification Criteria	Yes	No	N/A
Were common laboratory contaminants detected?		X	
Was professional judgment used to qualify data (if yes, list below)?	X		

The RPD between the primary and confirmation column for some explosives samples was above evaluation criteria. Qualification of data is shown in the table below.

Sample ID	Analysis	Analyte	RPD	Qual
G0044-19A	Explosives	RDX	44.8	J
PZ015-19A	Explosives	RDX	117.4	J
G0103-19A	Explosives	3-nitrotoluene	145.7	J

#### 10.0 Completeness

Completeness Criteria		No	N/A
Were any data rejected during the verification process?		X	
Were any samples lost, broken, or in any other manner in not verified?		X	
Were requested sample analyses performed, the correct analyte lists used, and correct sample preparation and analyses methods and units utilized?			