

## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-149116

AECOM Chemist: Jared DeSadier

Date Verified: 6/25/2021

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8260B, 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

Sample Identification #	Date Collected	Date Received	Matrix	Analysis
SHGW03-21A	5/26/2021	5/27/2021	Water	VOCs (8260B), Nitrate, Nitrite (353.2), MEE (RSK-175), Sulfate (9056A), DRO (8015C)
SHGW02-21A	5/26/2021	5/27/2021	Water	VOCs (8260B), Nitrate, Nitrite (353.2), MEE (RSK-175), Sulfate (9056A)
SHGW05-21A	5/26/2021	5/27/2021	Water	VOCs (8260B), Nitrate, Nitrite (353.2), MEE (RSK-175), Sulfate (9056A)
TB052621	5/26/2021	5/27/2021	Water	VOCs (8260B)
PZ020-21A	5/26/2021	5/27/2021	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), Methane (RSK-175), DOC (9060A), Sulfate (9056A), Alkalinity (2320B), Sulfide (9034)
G0077-21A	5/26/2021	5/27/2021	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), Methane (RSK-175), DOC (9060A), Sulfate (9056A), Alkalinity (2320B), Sulfide (9034)
G0078-21A	5/26/2021	5/27/2021	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), Methane (RSK-175), DOC (9060A), Sulfate (9056A), Alkalinity (2320B), Sulfide (9034)
PZ017R-21A	5/26/2021	5/27/2021	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), Methane (RSK-175), DOC (9060A), Sulfate (9056A), Alkalinity (2320B), Sulfide (9034)
PZ021-21A	5/26/2021	5/27/2021	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), Methane (RSK-175), DOC (9060A), Sulfate (9056A), Alkalinity (2320B), Sulfide (9034)
PZ007-21A	5/26/2021	5/27/2021	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), Methane (RSK-175), DOC (9060A), Sulfate (9056A), Alkalinity (2320B), Sulfide (9034)
G0024-21A	5/26/2021	5/27/2021	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), Methane (RSK-175), DOC (9060A), Sulfate (9056A), Alkalinity (2320B), Sulfide (9034)
G0017-21A	5/26/2021	5/27/2021	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), Methane (RSK-175), DOC (9060A), Sulfate (9056A), Alkalinity (2320B), Sulfide (9034)
G0045-21A	5/26/2021	5/27/2021	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), Methane (RSK-175), DOC (9060A), Sulfate (9056A), Alkalinity (2320B), Sulfide (9034)
G0086-21A	5/26/2021	5/27/2021	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), Methane (RSK-175), DOC (9060A), Sulfate (9056A), Alkalinity (2320B), Sulfide (9034)
PZ018-21A	5/26/2021	5/27/2021	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), Methane (RSK-175), DOC (9060A), Sulfate (9056A), Alkalinity (2320B), Sulfide (9034)

# CHAAP Data Verification

Laboratory and SDG#: TADenver 280-149116

AECOM Chemist: Jared DeSadier

Date Verified: 6/25/2021

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8260B, 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

## 1.0 Laboratory Case Narrative \ Cooler Receipt Form

Verification Criteria	Yes	No	N/A
Were any DoD QSM deviations noted in the laboratory case narrative?	X		
Were DoD QSM corrective actions followed if deviations were noted?	X		
Were any issues noted in the cooler receipt form?	X		

The case narrative indicated that some MS/MSD, LCS/LCSD and surrogate recoveries were outside evaluation criteria, and some analytes were detected in method blanks. These issues are discussed further in the ADR report. No qualification of data for blank contamination was required for nondetect analytes or those with a native concentration greater than five times the blank concentration. Qualification of data based on holding time was not required for dissolved gas samples as they were analyzed within the acceptable window for unpreserved samples.

The case narrative also indicated that the RPD between the primary and confirmation column for some explosives samples was above evaluation criteria and some dissolved gas VOAs were received with headspace greater than 6 mm. These issues are discussed further in Section 7.0. Some samples were further preserved by the lab upon receipt. No qualification was required.

The cooler receipt form indicated analyses were not marked on the COC for samples PZ017R-21A, PZ021-21A, and PZ020-21A. Samples were logged according to the received containers per the AECOM chemist and no qualification of data was required. The cooler receipt form also indicated that two samples were received outside of temperature criteria. This issue is discussed further in the ADR report.

No other issues were noted in the case narrative or cooler receipt form.

## 2.0 Sample Documentation

Verification Criteria	Yes	No
Were all samples documented correctly on the chain-of-custody (COC) and samples labels?		X
Were all sample identifications (IDs) documented correctly on sample labels?		X
Did samples listed on COCs match the sample labels?		X
Were samples relinquished properly on the COC?	X	

See Section 1.0.

## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-149116

AECOM Chemist: Jared DeSadier

Date Verified: 6/25/2021

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8260B, 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

### 3.0 Instrument Performance Check (Tuning)

Method 8260B Instrument Tuning Criteria (Filename)	MS1 2711.D		
Instrument:	VMS MS1		
Date of Tuning:	6/4/2021		
	Yes	No	N/A
Was instrument tuning completed prior to calibration?	X		
Were all samples analyzed under an acceptable 12 hour clock tune?	X		
Were ion relative abundances for each target mass within the required intensity limits listed in Table 4 of SW-846 Method 8260B?	X		

Method 8260B Instrument Tuning Criteria (Filename)	MS1 2743.D		
Instrument:	VMS MS1		
Date of Tuning:	6/5/2021		
	Yes	No	N/A
Was instrument tuning completed prior to calibration?	X		
Were all samples analyzed under an acceptable 12 hour clock tune?	X		
Were ion relative abundances for each target mass within the required intensity limits listed in Table 4 of SW-846 Method 8260B?	X		

### 4.0 Initial Calibration

Method 8260B Initial Calibration Criteria			
Instrument:	VMS MS1		
Date of Calibration:	6/4/2021		
	Yes	No	N/A
Option 1: RSD for each analyte $\leq 15\%$ ?	X		
Option 2: If linear least squares regression was used was the $r^2 \geq 0.99$ ?	X		
Option 3: If non-linear regression was used was the coefficient of determination $r^2 \geq 0.99$ ?			X
If non-linear regression was used were 6 points used for second order and 7 points for third order?			X

## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-149116

AECOM Chemist: Jared DeSadier

Date Verified: 6/25/2021

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8260B, 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

Method 8330A Initial Calibration Criteria			
Instrument:	CHHPLC_X3		
Date of Calibration:	3/2/2021		
	Yes	No	N/A
Was at least a five point calibration completed for all analytes prior to sample analysis and one option below?	X		
Option 1: RSD for each analyte $\leq 20\%$ ?	X		
Option 2: If linear least squares regression was used was the $r^2 \geq 0.99$ ?			X
Option 3: If non-linear regression was used was the coefficient of determination $r^2 \geq 0.99$ ?			X
If non-linear regression was used were 6 points used for second order and 7 points for third order?			X

Method 8330A Initial Calibration Criteria			
Instrument:	CHHPLC_X3		
Date of Calibration:	5/1/2021		
	Yes	No	N/A
Was at least a five point calibration completed for all analytes prior to sample analysis and one option below?	X		
Option 1: RSD for each analyte $\leq 20\%$ ?	X		
Option 2: If linear least squares regression was used was the $r^2 \geq 0.99$ ?			X
Option 3: If non-linear regression was used was the coefficient of determination $r^2 \geq 0.99$ ?			X
If non-linear regression was used were 6 points used for second order and 7 points for third order?			X

Method 8330A Initial Calibration Criteria			
Instrument:	CHHPLC_X5		
Date of Calibration:	5/1/2021		
	Yes	No	N/A
Was at least a five point calibration completed for all analytes prior to sample analysis and one option below?	X		
Option 1: RSD for each analyte $\leq 20\%$ ?	X		
Option 2: If linear least squares regression was used was the $r^2 \geq 0.99$ ?			X
Option 3: If non-linear regression was used was the coefficient of determination $r^2 \geq 0.99$ ?			X
If non-linear regression was used were 6 points used for second order and 7 points for third order?			X

## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-149116

AECOM Chemist: Jared DeSadier

Date Verified: 6/25/2021

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8260B, 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

Method 8330A Initial Calibration Criteria			
Instrument:	CHHPLC_X5		
Date of Calibration:	5/2/2021		
	Yes	No	N/A
Was at least a five point calibration completed for all analytes prior to sample analysis and one option below?	X		
Option 1: RSD for each analyte $\leq 20\%$ ?	X		
Option 2: If linear least squares regression was used was the $r^2 \geq 0.99$ ?			X
Option 3: If non-linear regression was used was the coefficient of determination $r^2 \geq 0.99$ ?			X
If non-linear regression was used were 6 points used for second order and 7 points for third order?			X

Method RSK-175 Initial Calibration Criteria			
Instrument:	VGC_J		
Date of Calibration:	6/24/2020		
	Yes	No	N/A
Was at least a five point calibration completed for all analytes prior to sample analysis and one option below?	X		
Option 1: RSD for each analyte $\leq 25\%$ ?	X		
Option 2: If linear least squares regression was used was the $r^2 \geq 0.99$ ?	X		
Option 3: If non-linear regression was used was the coefficient of determination $r^2 \geq 0.99$ ?			X
If non-linear regression was used were 6 points used for second order and 7 points for third order?			X

Method 9056A Initial Calibration Criteria			
Instrument:	WC IonChrom13		
Date of Calibration:	6/14/2021		
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \geq 0.99$ ?	X		

Method 9056A Initial Calibration Criteria			
Instrument:	24811		
Date of Calibration:	6/11/2021		
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \geq 0.99$ ?	X		

## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-149116

AECOM Chemist: Jared DeSadier

Date Verified: 6/25/2021

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8260B, 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

Method 350.1 Initial Calibration Criteria			
Instrument:	WC Alp 4		
Date of Calibration:	6/11/2021		
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \geq 0.99$ ?	X		

Method 353.2 Initial Calibration Criteria			
Instrument:	WC Alp 2		
Date of Calibration:	6/4/2021		
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \geq 0.99$ ?	X		

Method 351.2 Initial Calibration Criteria			
Instrument:	WC Astoria		
Date of Calibration:	6/2/2021		
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \geq 0.99$ ?	X		

Method 9060A Initial Calibration Criteria			
Instrument:	WC SHI4		
Date of Calibration:	6/8/2021		
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \geq 0.99$ ?	X		

### 5.0 Initial Calibration Verification [(ICV) Second Source]

Method 8260B ICV Criteria (Filename)			
Instrument:	MS1 2732.D		
Date of Initial Calibration Verification:	VMS_MS1		
	6/4/2021		
	Yes	No	N/A
Was the ICV analyzed after each calibration?	X		
Were all reported analytes within $\pm 20\%$ of true value?	X		

Method 8330A ICV Criteria (Filename)			
Instrument:	03020037.D		
Date of Initial Calibration Verification:	CHHPLC_X3		
	3/3/2021		
	Yes	No	N/A
Was the ICV analyzed after each calibration?	X		
Was the ICV for all analytes within $\pm 15\%$ of the true value?	X		

## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-149116

AECOM Chemist: Jared DeSadier

Date Verified: 6/25/2021

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8260B, 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

<b>Method 8330A ICV Criteria (Filename)</b>	<b>020-1401.D</b>		
<b>Instrument:</b>	<b>CHHPLC X3</b>		
<b>Date of Initial Calibration Verification:</b>	<b>5/1/2021</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the ICV analyzed after each calibration?	X		
Was the ICV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A ICV Criteria (Filename)</b>	<b>019-1401.D</b>		
<b>Instrument:</b>	<b>CHHPLC X5</b>		
<b>Date of Initial Calibration Verification:</b>	<b>5/2/2021</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the ICV analyzed after each calibration?	X		
Was the ICV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A ICV Criteria (Filename)</b>	<b>028-2301.D</b>		
<b>Instrument:</b>	<b>CHHPLC X5</b>		
<b>Date of Initial Calibration Verification:</b>	<b>5/2/2021</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the ICV analyzed after each calibration?	X		
Was the ICV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method RSK-175 ICV Criteria (Filename)</b>	<b>06242011.D</b>		
<b>Instrument:</b>	<b>VGC J</b>		
<b>Date of Initial Calibration Verification:</b>	<b>6/24/2020</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the ICV analyzed after each calibration?	X		
Was the ICV for all analytes within $\pm 25\%$ of the true value?	X		

<b>Method 9056A ICV</b>	<b>WC IonChrom13</b>		
<b>Date of Initial Calibration Verification:</b>	<b>6/14/2021</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm 10\%$ of the true value?	X		

<b>Method 9056A ICV</b>	<b>24811</b>		
<b>Date of Initial Calibration Verification:</b>	<b>6/11/2021</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm 10\%$ of the true value?	X		

## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-149116

AECOM Chemist: Jared DeSadier

Date Verified: 6/25/2021

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8260B, 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

Method 350.1 ICV Criteria	WC Alp 4		
Date of Initial Calibration Verification:	6/11/2021		
	Yes	No	N/A
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm 10\%$ of the true value?	X		

Method 353.2 ICV Criteria	WC Alp 2		
Date of Initial Calibration Verification:	6/4/2021		
	Yes	No	N/A
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm 10\%$ of the true value?	X		

Method 351.2 ICV Criteria	WC Astoria		
Date of Initial Calibration Verification:	6/2/2021		
	Yes	No	N/A
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm 10\%$ of the true value?	X		

Method 9060A ICV Criteria	WC SHI4		
Date of Initial Calibration Verification:	6/8/2021		
	Yes	No	N/A
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm 10\%$ of the true value?	X		

### 6.0 Continuing Calibration Verification (CCV)

Method 8260B Beginning CCV Criteria (Filename)	MS1 2744.D		
Method 8260B Ending CCV Criteria (Filename)	MS1 2773.D		
Instrument:	VMS MS1		
Date of Calibration Verification:	6/5/2021		
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 12 hours of analysis time?	X		
Were all reported analytes and surrogates within $\pm 20\%$ of true value?	X		
Were all reported analytes and surrogates within $\pm 50\%$ of true value for the end of analytical batch CCV?	X		



## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-149116

AECOM Chemist: Jared DeSadier

Date Verified: 6/25/2021

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8260B, 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

<b>Method 8330A CCV Criteria (Filename)</b>	<b>06030012-3.D</b>		
<b>Instrument:</b>	<b>CHHPLC X3</b>		
<b>Date of Calibration Verification:</b>	<b>6/3/2021</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A CCV Criteria (Filename)</b>	<b>06030024-6.D</b>		
<b>Instrument:</b>	<b>CHHPLC X3</b>		
<b>Date of Calibration Verification:</b>	<b>6/3/2021</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A CCV Criteria (Filename)</b>	<b>06030038-9.D</b>		
<b>Instrument:</b>	<b>CHHPLC X3</b>		
<b>Date of Calibration Verification:</b>	<b>6/4/2021</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A CCV Criteria (Filename)</b>	<b>06040007-9.D</b>		
<b>Instrument:</b>	<b>CHHPLC X3</b>		
<b>Date of Calibration Verification:</b>	<b>6/4/2021</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-149116

AECOM Chemist: Jared DeSadier

Date Verified: 6/25/2021

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8260B, 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

<b>Method 8330A CCV Criteria (Filename)</b>	<b>06040021-3.D</b>		
<b>Instrument:</b>	<b>CHHPLC X3</b>		
<b>Date of Calibration Verification:</b>	<b>6/4/2021</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A CCV Criteria (Filename)</b>	<b>06040034-6.D</b>		
<b>Instrument:</b>	<b>CHHPLC X3</b>		
<b>Date of Calibration Verification:</b>	<b>6/5/2021</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A CCV Criteria (Filename)</b>	<b>06040046-8.D</b>		
<b>Instrument:</b>	<b>CHHPLC X3</b>		
<b>Date of Calibration Verification:</b>	<b>6/5/2021</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A CCV Criteria (Filename)</b>	<b>06040007.D</b>		
<b>Instrument:</b>	<b>CHHPLC X5</b>		
<b>Date of Calibration Verification:</b>	<b>6/4/2021</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-149116

AECOM Chemist: Jared DeSadier

Date Verified: 6/25/2021

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8260B, 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

<b>Method 8330A CCV Criteria (Filename)</b>	<b>06040015.D</b>		
<b>Instrument:</b>	<b>CHHPLC X5</b>		
<b>Date of Calibration Verification:</b>	<b>6/4/2021</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A CCV Criteria (Filename)</b>	<b>06050007-8.D</b>		
<b>Instrument:</b>	<b>CHHPLC X5</b>		
<b>Date of Calibration Verification:</b>	<b>6/5/2021</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A CCV Criteria (Filename)</b>	<b>06050020-1.D</b>		
<b>Instrument:</b>	<b>CHHPLC X5</b>		
<b>Date of Calibration Verification:</b>	<b>6/5/2021</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A CCV Criteria (Filename)</b>	<b>06050026-7.D</b>		
<b>Instrument:</b>	<b>CHHPLC X5</b>		
<b>Date of Calibration Verification:</b>	<b>6/5-6/2021</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-149116

AECOM Chemist: Jared DeSadier

Date Verified: 6/25/2021

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8260B, 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

Method RSK-175 CCVRT Criteria (Filename)	004F0401.D		
Instrument:	VGC J		
Date of Calibration Verification:	6/1/2021		
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 25\%$ of the true value?	X		

Method RSK-175 CCV Criteria (Filename)	021F2101.D		
Instrument:	VGC J		
Date of Calibration Verification:	6/1/2021		
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 25\%$ of the true value?	X		

Method RSK-175 CCV Criteria (Filename)	035F3501.D		
Instrument:	VGC J		
Date of Calibration Verification:	6/1/2021		
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 25\%$ of the true value?	X		

Method RSK-175 CCV Criteria (Filename)	054F5401.D		
Instrument:	VGC J		
Date of Calibration Verification:	6/1/2021		
	Yes	No	N/A
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 25\%$ of the true value?	X		

## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-149116

AECOM Chemist: Jared DeSadier

Date Verified: 6/25/2021

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8260B, 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

Method RSK-175 CCV Criteria (Filename)		066F6601.D		
Instrument:		VGC J		
Date of Calibration Verification:		6/2/2021		
		Yes	No	N/A
Was the CCV analyzed daily before sample analysis?		X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?		X		
Was the CCV for all analytes within $\pm 25\%$ of the true value?		X		

Method 9056A, Instrument: WC_IonChrom13, All CCVs on 6/14/2021		Yes	No
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?		X	
Were the CCVs for all analytes within $\pm 10\%$ of the true value?		X	

Method 9056A, Instrument: WC_IonChrom13, All CCVs on 6/16/2021		Yes	No
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?		X	
Were the CCVs for all analytes within $\pm 10\%$ of the true value?		X	

Method 9056A, Instrument: 24811, All CCVs on 6/16/2021		Yes	No
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?		X	
Were the CCVs for all analytes within $\pm 10\%$ of the true value?		X	

Method 350.1, Instrument: WC_Alp 4, All CCVs on 6/11/2021		Yes	No
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?		X	
Were the CCVs for all analytes within $\pm 10\%$ of the true value?		X	

Method 353.2, Instrument: WC_Alp 2, All CCVs on 6/4/2021		Yes	No
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?		X	
Were the CCVs for all analytes within $\pm 10\%$ of the true value?		X	

Method 351.2, Instrument: WC_Astoria, All CCVs on 6/2/2021		Yes	No
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?		X	
Were the CCVs for all analytes within $\pm 10\%$ of the true value?		X	

Method 9060A, Instrument: WC_SHI4, All CCVs on 6/8/2021		Yes	No
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?		X	
Were the CCVs for all analytes within $\pm 10\%$ of the true value?		X	

### 7.0 Sensitivity

Sensitivity Criteria	Yes	No	N/A
Was the laboratory sensitivity consistent with project (QAPP) requirements?	X		
Did all analytes meet sensitivity requirements?	X		

## CHAAP Data Verification

Laboratory and SDG#: TADenver 280-149116

AECOM Chemist: Jared DeSadier

Date Verified: 6/25/2021

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8260B, 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

### 8.0 Internal Standard (IS) Recoveries

Method 8260B Criteria	Yes	No	N/A
Were internal standards spiked for all samples and standards?	X		
Were internal standard areas within -50% to + 100% of the ICAL midpoint standard area?	X		
Were retention time $\pm$ 30 seconds from the retention time of the midpoint standard of the ICAL?	X		

### 9.0 Additional Qualifications

Additional Qualification Criteria	Yes	No	N/A
Were common laboratory contaminants detected?		X	
Was professional judgment used to qualify data (if yes, list below)?	X		

The RPD between the primary and confirmation column for some explosives samples was above evaluation criteria. Qualification of data is shown in the table below; results were reported from primary column unless otherwise noted.

Sample ID	Analysis	Analyte	RPD	Qual
G0017-21A	Explosives	2-amino-4,6-dinitrotoluene	154.4	J
G0086-21A	Explosives	RDX	45.0	J
PZ018-21A	Explosives	4-amino-2,6-dinitrotoluene	65.0	J
G0077-21A	Explosives	RDX	65.2	J
PZ017R-21A	Explosives	HMX	91.6	J
PZ017R-21A	Explosives	2,6-dinitrotoluene	81.4	J
PZ021-21A	Explosives	HMX	80.3	J
PZ021-21A	Explosives	2,6-dinitrotoluene	88.1	J
PZ020-21A	Explosives	RDX	50.2	J

Sample G0045-21A was received with headspace greater than 6 mm in its sample VOAs. Qualification of data is shown in the table below.

Sample ID	Analysis	Analyte	Qual
G0045-21A	RSK-175	Methane	J

# CHAAP Data Verification

Laboratory and SDG#: TADenver 280-149116

AECOM Chemist: Jared DeSadier

Date Verified: 6/25/2021

AECOM ITR: Jeff Aust

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2018)

Applicable Analytical Methods: 8260B, 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

## 10.0 Completeness

Completeness Criteria	Yes	No	N/A
Were any data rejected during the verification process?		X	
Were any samples lost, broken, or in any other manner in not verified?		X	
Were requested sample analyses performed, the correct analyte lists used, and correct sample preparation and analyses methods and units utilized?	X		