

## CHAAP Data Verification

Laboratory and SDG#: Eurofins 280-162320

AECOM Chemist: D. Casagrande

Date Verified: 7/20/2022

AECOM ITR: S. Louie

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2019)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

Sample Identification #	Date Collected	Date Received	Matrix	Analysis
G0080-22A	5/11/2022	5/13/2022	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), Methane (RSK-175), DOC (9060A), Sulfate (9056A), Alkalinity (2320B), Sulfide (9034)
G0082-22A	5/11/2022	5/13/2022	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), Methane (RSK-175), DOC (9060A), Sulfate (9056A), Alkalinity (2320B), Sulfide (9034)
G0083-22A	5/11/2022	5/13/2022	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), Methane (RSK-175), DOC (9060A), Sulfate (9056A), Alkalinity (2320B), Sulfide (9034)
G0075-22A	5/12/2022	5/13/2022	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), Methane (RSK-175), DOC (9060A), Sulfate (9056A), Alkalinity (2320B), Sulfide (9034)
G0076-22A	5/12/2022	5/13/2022	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), Methane (RSK-175), DOC (9060A), Sulfate (9056A), Alkalinity (2320B), Sulfide (9034)
G0091-22A	5/12/2022	5/13/2022	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), Methane (RSK-175), DOC (9060A), Sulfate (9056A), Alkalinity (2320B), Sulfide (9034)
G0070-22A	5/12/2022	5/13/2022	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), Methane (RSK-175), DOC (9060A), Sulfate (9056A), Alkalinity (2320B), Sulfide (9034)
NW060-22A	5/12/2022	5/13/2022	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), Methane (RSK-175), DOC (9060A), Sulfate (9056A), Alkalinity (2320B), Sulfide (9034)
NW061-22A	5/12/2022	5/13/2022	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), Methane (RSK-175), DOC (9060A), Sulfate (9056A), Alkalinity (2320B), Sulfide (9034)
NW062-22A	5/12/2022	5/13/2022	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), Methane (RSK-175), DOC (9060A), Sulfate (9056A), Alkalinity (2320B), Sulfide (9034)
G0079-22A	5/12/2022	5/13/2022	Water	Explosives (8330A), Nitrate, Nitrite (353.2), Ammonia (350.1), TKN (351.2), Methane (RSK-175), DOC (9060A), Sulfate (9056A), Alkalinity (2320B), Sulfide (9034)

# CHAAP Data Verification

Laboratory and SDG#: Eurofins 280-162320

AECOM Chemist: D. Casagrande

Date Verified: 7/20/2022

AECOM ITR: S. Louie

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2019)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

## 1.0 Laboratory Case Narrative \ Cooler Receipt Form

Verification Criteria	Yes	No	N/A
Were any DoD QSM deviations noted in the laboratory case narrative?	X		
Were DoD QSM corrective actions followed if deviations were noted?	X		
Were any issues noted in the cooler receipt form?		X	

*Validator comments in italics.*

Method SW8330A:

3-Nitrotoluene failed the recovery criteria low for the MSD of sample NW062-22AMSD (280-162320-11) in batch 280-575546. 3-Nitrotoluene failed the recovery criteria low for the MS of sample G0070-22AMS (280-162320-7) in batch 280-575414. 3-Nitrotoluene failed the recovery criteria low for the MSD of sample G0070-22AMSD (280-162320-7) in batch 280-575414. *3-Nitrotoluene was not detected in samples NW062-22A or G007-22A, therefore these results are qualified (UJ). These issues are further discussed in the ADR report.*

The %RPD between the primary and confirmation column exceeded 40% for 2-Amino-4,6-dinitrotoluene, HMX and RDX for the following samples: G0082-22A (280-162320-2), G0075-22A (280-162320-4) and G0091-22A (280-162320-6) in preparation batch 280-575370 and analytical batch 280-575414 for method 8330. The results from both columns has been qualified and reported in accordance with the laboratory's QAS. *This issue is discussed further in Section 7.0.*

Method 350.1:

Ammonia failed the recovery criteria high for the MSD of sample G0070-22AMSD (280-162320-7) in batch 280-576196. Ammonia exceeded the RPD limit. *Ammonia was detected in sample G007-22A, therefore this result is qualified (J). This issue is further discussed in the ADR report.*

Method 351.2:

Nitrogen, Total Kjeldahl failed the recovery criteria high for the MS of sample G0070-22AMS (280-162320-7) in batch 280-576217. Nitrogen, Total Kjeldahl failed the recovery criteria high for the MSD of sample G0070-22AMSD (280-162320-7) in batch 280-576217. *No data are qualified or affected, since TKN is not detected in the associated sample. This issue is further discussed in the ADR report.*

No other issues were noted in the case narrative or cooler receipt form for all other methods.

## 2.0 Sample Documentation

Verification Criteria	Yes	No
Were all samples documented correctly on the chain-of-custody (COC) and samples labels?	X	
Were all sample identifications (IDs) documented correctly on sample labels?	X	
Did samples listed on COCs match the sample labels?	X	
Were samples relinquished properly on the COC?	X	

## CHAAP Data Verification

Laboratory and SDG#: Eurofins 280-162320

AECOM Chemist: D. Casagrande

Date Verified: 7/20/2022

AECOM ITR: S. Louie

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2019)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

### 3.0 Initial Calibration

Method 8330A Initial Calibration Criteria			
Instrument:	CHHPLC X3		
Date of Calibration:	1/4/2022		
	Yes	No	N/A
Was at least a five point calibration completed for all analytes prior to sample analysis and one option below?	X		
Option 1: RSD for each analyte $\leq 20\%$ ?	X		
Option 2: If linear least squares regression was used was the $r^2 \geq 0.99$ ?			X
Option 3: If non-linear regression was used was the coefficient of determination $r^2 \geq 0.99$ ?			X
If non-linear regression was used were 6 points used for second order and 7 points for third order?			X

%RSDs were met for all target analytes.

Method 8330A Initial Calibration Criteria			
Instrument:	CHHPLC X3		
Date of Calibration:	1/5/2022		
	Yes	No	N/A
Was at least a five point calibration completed for all analytes prior to sample analysis and one option below?	X		
Option 1: RSD for each analyte $\leq 20\%$ ?	X		
Option 2: If linear least squares regression was used was the $r^2 \geq 0.99$ ?			X
Option 3: If non-linear regression was used was the coefficient of determination $r^2 \geq 0.99$ ?			X
If non-linear regression was used were 6 points used for second order and 7 points for third order?			X

Method 8330A Initial Calibration Criteria			
Instrument:	CHHPLC X5		
Date of Calibration:	3/2/2022		
	Yes	No	N/A
Was at least a five point calibration completed for all analytes prior to sample analysis and one option below?	X		
Option 1: RSD for each analyte $\leq 20\%$ ?	X		
Option 2: If linear least squares regression was used was the $r^2 \geq 0.99$ ?	X		
Option 3: If non-linear regression was used was the coefficient of determination $r^2 \geq 0.99$ ?			X
If non-linear regression was used were 6 points used for second order and 7 points for third order?			X

## CHAAP Data Verification

Laboratory and SDG#: Eurofins 280-162320

AECOM Chemist: D. Casagrande

Date Verified: 7/20/2022

AECOM ITR: S. Louie

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2019)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

Method 8330A Initial Calibration Criteria			
Instrument:	CHHPLC_X5		
Date of Calibration:	3/3/2022		
	Yes	No	N/A
Was at least a five point calibration completed for all analytes prior to sample analysis and one option below?	X		
Option 1: RSD for each analyte $\leq 20\%$ ?	X		
Option 2: If linear least squares regression was used was the $r^2 \geq 0.99$ ?			X
Option 3: If non-linear regression was used was the coefficient of determination $r^2 \geq 0.99$ ?			X
If non-linear regression was used were 6 points used for second order and 7 points for third order?			X

Method RSK-175 Initial Calibration Criteria			
Instrument:	VGC_J		
Date of Calibration:	9/24/2021		
	Yes	No	N/A
Was at least a five point calibration completed for all analytes prior to sample analysis and one option below?	X		
Option 1: RSD for each analyte $\leq 25\%$ ?			X
Option 2: If linear least squares regression was used was the $r^2 \geq 0.99$ ?	X		
Option 3: If non-linear regression was used was the coefficient of determination $r^2 \geq 0.99$ ?			X
If non-linear regression was used were 6 points used for second order and 7 points for third order?			X

%RSD was not provided for methane; however, the r2 was met.

Method 9056A Initial Calibration Criteria			
Instrument:	WC IonChrom11		
Date of Calibration:	5/26/2022		
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \geq 0.99$ ?	X		

Method 9056A Initial Calibration Criteria			
Instrument:	WC IonChrom11		
Date of Calibration:	5/24/2022		
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \geq 0.99$ ?	X		

## CHAAP Data Verification

Laboratory and SDG#: Eurofins 280-162320

AECOM Chemist: D. Casagrande

Date Verified: 7/20/2022

AECOM ITR: S. Louie

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2019)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

Method 350.1 Initial Calibration Criteria			
Instrument:	WC Alp 4		
Date of Calibration:	5/27/2022		
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \geq 0.99$ ?	X		

Method 350.1 Initial Calibration Criteria			
Instrument:	WC_SKALAR_01		
Date of Calibration:	5/25/2022		
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \geq 0.99$ ?	X		

Method 353.2 Initial Calibration Criteria			
Instrument:	WC Alp 2		
Date of Calibration:	5/24/2022		
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \geq 0.99$ ?	X		

Method 351.2 Initial Calibration Criteria			
Instrument:	WC GAL1		
Date of Calibration:	5/25/2022		
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \geq 0.99$ ?	X		

Method 9060A Initial Calibration Criteria			
Instrument:	WC_SHI5		
Date of Calibration:	12/10/2021		
	Yes	No	N/A
Was a minimum of three standards and a calibration blank used for ICAL?	X		
Was $r^2 \geq 0.99$ ?	X		

## CHAAP Data Verification

Laboratory and SDG#: Eurofins 280-162320

AECOM Chemist: D. Casagrande

Date Verified: 7/20/2022

AECOM ITR: S. Louie

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2019)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

### 4.0 Initial Calibration Verification [(ICV) Second Source]

<b>Method 8330A ICV Criteria (Filename)</b>	<b>280-562503/20</b>		
<b>Instrument:</b>	<b>CHHPLC X3</b>		
<b>Date of Initial Calibration Verification:</b>	<b>1/4/2022</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the ICV analyzed after each calibration?	X		
Was the ICV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A ICV Criteria (Filename)</b>	<b>280-562503/38</b>		
<b>Instrument:</b>	<b>CHHPLC X3</b>		
<b>Date of Initial Calibration Verification:</b>	<b>1/5/2022</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the ICV analyzed after each calibration?	X		
Was the ICV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A ICV Criteria (Filename)</b>	<b>280-567560/19</b>		
<b>Instrument:</b>	<b>CHHPLC X5</b>		
<b>Date of Initial Calibration Verification:</b>	<b>3/3/2022</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the ICV analyzed after each calibration?	X		
Was the ICV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A ICV Criteria (Filename)</b>	<b>280-567560/28</b>		
<b>Instrument:</b>	<b>CHHPLC X5</b>		
<b>Date of Initial Calibration Verification:</b>	<b>3/3/2022</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the ICV analyzed after each calibration?	X		
Was the ICV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method RSK-175 ICV Criteria (Filename)</b>	<b>280-550959/13</b>		
<b>Instrument:</b>	<b>VGC J</b>		
<b>Date of Initial Calibration Verification:</b>	<b>9/24/2021</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the ICV analyzed after each calibration?	X		
Was the ICV for all analytes within $\pm 25\%$ of the true value?	X		

## CHAAP Data Verification

Laboratory and SDG#: Eurofins 280-162320

AECOM Chemist: D. Casagrande

Date Verified: 7/20/2022

AECOM ITR: S. Louie

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2019)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

Method 9056A ICV	WC IonChrom11		
Date of Initial Calibration Verification:	5/26/2022		
	Yes	No	N/A
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm 10\%$ of the true value?	X		

Method 350.1 ICV Criteria	WC Alp 4		
Date of Initial Calibration Verification:	5/27/2022		
	Yes	No	N/A
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm 10\%$ of the true value?	X		

Method 350.1 ICV Criteria	WC SKALAR_01		
Date of Initial Calibration Verification:	5/25/2022		
	Yes	No	N/A
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm 10\%$ of the true value?	X		

Method 353.2 ICV Criteria	WC Alp 2		
Date of Initial Calibration Verification:	5/24/2022		
	Yes	No	N/A
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm 10\%$ of the true value?	X		

Method 351.2 ICV Criteria	WC GAL1		
Date of Initial Calibration Verification:	5/25/2022		
	Yes	No	N/A
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm 10\%$ of the true value?	X		

Method 9060A ICV Criteria	WC SHI5		
Date of Initial Calibration Verification:	5/24/2022		
	Yes	No	N/A
Was the ICV analyzed after each ICAL, prior to the beginning of a sample analysis?	X		
Was the ICV for all analytes within $\pm 10\%$ of the true value?	X		

## CHAAP Data Verification

Laboratory and SDG#: Eurofins 280-162320

AECOM Chemist: D. Casagrande

Date Verified: 7/20/2022

AECOM ITR: S. Louie

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2019)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

### 5.0 Continuing Calibration Verification (CCV)

<b>Method 8330A CCV Criteria (Filename)</b>	<b>280-575414/32-33</b>		
<b>Instrument:</b>	<b>CHHPLC_X3</b>		
<b>Date of Calibration Verification:</b>	<b>5/19/2022</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A CCV Criteria (Filename)</b>	<b>280-575414/44-45</b>		
<b>Instrument:</b>	<b>CHHPLC_X3</b>		
<b>Date of Calibration Verification:</b>	<b>5/19/2022</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A CCV Criteria (Filename)</b>	<b>280-575414/56-57</b>		
<b>Instrument:</b>	<b>CHHPLC_X3</b>		
<b>Date of Calibration Verification:</b>	<b>5/19/2022</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

CCV criteria were met for all target analytes.

<b>Method 8330A CCV Criteria (Filename)</b>	<b>280-575414/66-67</b>		
<b>Instrument:</b>	<b>CHHPLC_X3</b>		
<b>Date of Calibration Verification:</b>	<b>5/19/2022</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

CCV criteria were met for all target analytes.



## CHAAP Data Verification

Laboratory and SDG#: Eurofins 280-162320

AECOM Chemist: D. Casagrande

Date Verified: 7/20/2022

AECOM ITR: S. Louie

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2019)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

<b>Method 8330A CCV Criteria (Filename)</b>	<b>280-575546/20-21</b>		
<b>Instrument:</b>	<b>CHHPLC X5</b>		
<b>Date of Calibration Verification:</b>	<b>5/19/2022</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A CCV Criteria (Filename)</b>	<b>280-575546/30-31</b>		
<b>Instrument:</b>	<b>CHHPLC X5</b>		
<b>Date of Calibration Verification:</b>	<b>5/19/2022</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method 8330A CCV Criteria (Filename)</b>	<b>280-575546/47-48</b>		
<b>Instrument:</b>	<b>CHHPLC X5</b>		
<b>Date of Calibration Verification:</b>	<b>5/19/2022</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 15\%$ of the true value?	X		

<b>Method RSK-175 CCVRT Criteria (Filename)</b>	<b>280-575858/2</b>		
<b>Instrument:</b>	<b>VGC J</b>		
<b>Date of Calibration Verification:</b>	<b>5/23/2022</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 25\%$ of the true value?	X		

## CHAAP Data Verification

Laboratory and SDG#: Eurofins 280-162320

AECOM Chemist: D. Casagrande

Date Verified: 7/20/2022

AECOM ITR: S. Louie

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2019)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

<b>Method RSK-175 CCV Criteria (Filename)</b>	<b>280-575858/17</b>		
<b>Instrument:</b>	<b>VGC J</b>		
<b>Date of Calibration Verification:</b>	<b>5/23/2022</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 25\%$ of the true value?	X		

<b>Method RSK-175 CCV Criteria (Filename)</b>	<b>280-575858/31</b>		
<b>Instrument:</b>	<b>VGC J</b>		
<b>Date of Calibration Verification:</b>	<b>5/23/2022</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 25\%$ of the true value?	X		

<b>Method RSK-175 CCVRT Criteria (Filename)</b>	<b>280-575860/2</b>		
<b>Instrument:</b>	<b>VGC J</b>		
<b>Date of Calibration Verification:</b>	<b>5/24/2022</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 25\%$ of the true value?	X		

<b>Method RSK-175 CCV Criteria (Filename)</b>	<b>280-575860/17</b>		
<b>Instrument:</b>	<b>VGC J</b>		
<b>Date of Calibration Verification:</b>	<b>5/24/2022</b>		
	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the CCV analyzed daily before sample analysis?	X		
Was the CCV analyzed every 10 field samples and at the end of the analysis sequence?	X		
Was the CCV for all analytes within $\pm 25\%$ of the true value?	X		

<b>Method 9056A, Instrument: WC IonChrom11, All CCVs on 5/26/2022</b>	<b>Yes</b>	<b>No</b>
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X	
Were the CCVs for all analytes within $\pm 10\%$ of the true value?	X	

<b>Method 9056A, Instrument: WC IonChrom11, All CCVs on 5/27/2022</b>	<b>Yes</b>	<b>No</b>
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X	
Were the CCVs for all analytes within $\pm 10\%$ of the true value?	X	

# CHAAP Data Verification

Laboratory and SDG#: Eurofins 280-162320

AECOM Chemist: D. Casagrande

Date Verified: 7/20/2022

AECOM ITR: S. Louie

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2019)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

<b>Method 350.1, Instrument: WC Alp 4, All CCVs on 5/27/2022</b>	<b>Yes</b>	<b>No</b>
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X	
Were the CCVs for all analytes within $\pm 10\%$ of the true value?	X	

<b>Method 350.1, Instrument: WC SKALAR 01, All CCVs on 5/25/2022</b>	<b>Yes</b>	<b>No</b>
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X	
Were the CCVs for all analytes within $\pm 10\%$ of the true value?	X	

<b>Method 353.2, Instrument: WC Alp 2, All CCVs on 5/24/2022</b>	<b>Yes</b>	<b>No</b>
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X	
Were the CCVs for all analytes within $\pm 10\%$ of the true value?	X	

<b>Method 351.2, Instrument: WC GAL1, All CCVs on 5/25/2022</b>	<b>Yes</b>	<b>No</b>
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X	
Were the CCVs for all analytes within $\pm 10\%$ of the true value?	X	

<b>Method 9060A, Instrument: WC SHI5, All CCVs on 5/24/2022 and 5/25/2022</b>	<b>Yes</b>	<b>No</b>
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X	
Were the CCVs for all analytes within $\pm 10\%$ of the true value?	X	

<b>Method 2320B, Instrument: WC AT4, All CCVs on 5/19/2022 and 5/20/2022</b>	<b>Yes</b>	<b>No</b>
Was a CCV analyzed after every 10 field samples and at the end of the analysis sequence?	X	
Were the CCVs for all analytes within $\pm 10\%$ of the true value?	X	

## 6.0 Sensitivity

<b>Sensitivity Criteria</b>	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Was the laboratory sensitivity consistent with project (QAPP) requirements?	X		
Did all analytes meet sensitivity requirements?	X		

## 7.0 Additional Qualifications

<b>Additional Qualification Criteria</b>	<b>Yes</b>	<b>No</b>	<b>N/A</b>
Were common laboratory contaminants detected?		X	
Was professional judgment used to qualify data (if yes, list below)?	X		

The RPD between the primary and confirmation column for some explosives were above evaluation criteria. Qualification of data is shown in the table below; results were reported from the primary column unless otherwise noted.

## CHAAP Data Verification

Laboratory and SDG#: Eurofins 280-162320

AECOM Chemist: D. Casagrande

Date Verified: 7/20/2022

AECOM ITR: S. Louie

Guidance: DoD QSM Version 5.1 (January 2017)

Applicable QAPP: Cornhusker Army Ammunition Plant QAPP (Brice and AECOM, October 2019)

Applicable Analytical Methods: 8330A, 353.2, 350.1, 351.2, RSK-175, 9060A, 2320B, 9056A, 9034

Sample ID	Analysis	Analyte	RPD	Qual
G0082-22A	Explosives	2-amino-4,6-dinitrotoluene	66.2	J
G0075-22A	Explosives	HMX	73.7	J
G0091-22A	Explosives	2-amino-4,6-dinitrotoluene	57.6	J
		RDX	52.6	J

### 8.0 Completeness

Completeness Criteria	Yes	No	N/A
Were any data rejected during the verification process?		X	
Were any samples lost, broken, or in any other manner in not verified?		X	
Were requested sample analyses performed, the correct analyte lists used, and correct sample preparation and analyses methods and units utilized?	X		